

Certificate of Analysis

Product name	AMPLIRUN® TOTAL MTB RIF RESISTANT CONTROL (SPUTUM)
Catalogue Number	MBTC014
Lot Number	23MBTC014001-C
Expiry date	2025-09-14

Kit contents	Lot Number	Expiry date
VIRCELL TOTAL MTB RIF RESISTANT (531) CONTROL (SPUTUM)	23MBTC014101-C	2025-09-14

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	20000-50000	46000
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION
Identity test oligonucleotides <i>IS6110f</i> TCAAGGAGCACATCAGCC <i>IS6110r</i> TCACGGTTCAGGGTTAGC <i>Sanjuan-Jimenez R., et al. (2013). PLoS One, 8(3): e58353</i>		


VIRCELL TOTAL MTB RIF RESISTANT (526) CONTROL (SPUTUM)	23MBTC014201-C	2025-09-14
--	----------------	------------

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	20000-50000	35000
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION
Identity test oligonucleotides <i>IS6110f</i> TCAAGGAGCACATCAGCC <i>IS6110r</i> TCACGGTTCAGGGTTAGC <i>Sanjuan-Jimenez R., et al. (2013). PLoS One, 8(3): e58353</i>		

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Ramírez Galdeano, Jose Francisco	03/04/2023	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com