

# Certificate of Analysis

Product name	AMPLIRUN® TOTAL CMV CONTROL (PLASMA)
Catalogue Number	MBTC018-R
Lot Number	25MBTC018001-R
Expiry date	2027-03-02

Kit contents	Lot Number	Expiry date
VIRCELL TOTAL CMV CONTROL (PLASMA)	25MBTC018101-R	2027-03-02

  

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	4000-10000	6000
QUANTIFICATION TEST (IU/vial)	2350 - 5875	3400
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION

  

<b>Identity test oligonucleotides</b> UL75 F CCTTGCGTGTCTCGTATTCTAGC UL75 R GCCTCATCATCAACCAACGGACAG Onishi, Y., et al. (2006) Tohoku J Exp Med. 210(2):125-35
---

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	28/01/2025	