

Certificate of Analysis

Product name AMPLIRUN® TOTAL CMV CONTROL (PLASMA)
Catalogue Number MBTC018-R
Lot Number 24MBTC018001-R
Expiry date 2027-03-02

Kit contents	Lot Number	Expiry date
VIRCELL TOTAL CMV CONTROL (PLASMA)	24MBTC018101-R	2027-03-02

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	4000-10000	6000
QUANTIFICATION TEST (IU/vial)	2350 - 5875	3400
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION
Identity test oligonucleotides UL75 F CCTTGCGTGTCTCGTATTCTAGC UL75 R GCCTCATCATCAACCAACGGACAG Onishi, Y., et al. (2006) Tohoku J Exp Med. 210(2):125-35		

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	11/09/2024	