

# Certificate of Analysis

Product name	AMPLIRUN® TOTAL CLOSTRIDIUM DIFFICILE RT027 CONTROL (STOOL)
Catalogue Number	MBTC026-R
Lot Number	23MBTC026003-R
Expiry date	2025-12-29

Kit contents	Lot Number	Expiry date
VIRCELL TOTAL CDI027 CONTROL (STOOL)	23MBTC026103-R	2025-12-29

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/vial)	300000-750000	750000
IDENTITY TEST	SPECIFIC AMPLIFICATION	SPECIFIC AMPLIFICATION

**Identity test oligonucleotides**

CdtB-F AAAAGCTTCAGGTTCTTTTGACAAG  
CdtB-R TGATCAGTAGAGGCATGTTTCATT  
Wroblewski D., et al. (2009). J Clin Microbiol. 47(7):2142-8

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	14/02/2024	